



A comparison of global indices between the Medmont Automated Perimeter and the Humphrey Field Analyzer

John Landers, Alok Sharma, Ivan Goldberg, et al.

Br J Ophthalmol 2007 91: 1285-1287 originally published online March 27, 2007

doi: 10.1136/bjo.2007.114926

Updated information and services can be found at:

<http://bjo.bmj.com/content/91/10/1285.full.html>

These include:

References

This article cites 14 articles, 5 of which can be accessed free at:
<http://bjo.bmj.com/content/91/10/1285.full.html#ref-list-1>

Article cited in:

<http://bjo.bmj.com/content/91/10/1285.full.html#related-urls>

Open Access

This paper is freely available online under the BMJ Journals unlocked scheme, see <http://bjo.bmj.com/info/unlocked.dtl>

Email alerting service

Receive free email alerts when new articles cite this article. Sign up in the box at the top right corner of the online article.

Topic collections

Articles on similar topics can be found in the following collections

[Angle](#) (676 articles)
[Intraocular pressure](#) (674 articles)
[Unlocked](#) (190 articles)
[Glaucoma](#) (696 articles)

Notes

To order reprints of this article go to:

<http://bjo.bmj.com/cgi/reprintform>

To subscribe to *British Journal of Ophthalmology* go to:

<http://bjo.bmj.com/subscriptions>

SCIENTIFIC REPORT

A comparison of global indices between the Medmont Automated Perimeter and the Humphrey Field Analyzer

John Landers, Alok Sharma, Ivan Goldberg, Stuart Graham



This paper is freely available online under the BMJ Journals unlocked scheme, see <http://bjo.bmj.com/info/unlocked.dtl>

Br J Ophthalmol 2007;**91**:1285–1287. doi: 10.1136/bjo.2007.114926

Background: Two commonly used perimeters in Australia are the Humphrey Field Analyzer II (HFA) and the Medmont Automated Perimeter (MAP). Each device describes the visual field in terms of numerical values called global indices; however, these values are not interchangeable between devices. This study was designed to directly compare the global indices of HFA and MAP visual fields.

Methods: 63 subjects who had suspected glaucoma, ocular hypertension or glaucoma, or were normal controls were recruited selectively. Each patient was tested with the MAP and HFA. Global indices were then compared between tests. These included mean deviation (MD) and pattern standard deviation (PSD) from the HFA and average defect (AD) and pattern defect (PD) from the MAP.

Results: The MD and PSD results were strongly correlated with the AD and PD results, respectively. The relationship between them could be described in terms of two polynomial equations: $AD = 0.94 + 1.31(MD) + 0.02(MD)^2$ and $PD = 2.21(PSD) - 0.05(PSD)^2 - 0.006$. These non-linear relationships may be the result of differences in testing method (test stimulus spectrum, number of testing locations or background luminance) or differences in the way each global index was calculated.

Conclusion: The AD and PD results obtained from the MAP may be substituted for the MD and PSD results from the HFA after appropriate conversion.

During the past 30 years, automated perimetry has become the method to detect visual field loss.^{1–3} Two devices are the Medmont Automated Perimeter (MAP) and the Humphrey Field Analyzer II (HFA), which has become the standard against which others are compared.

The visual field may be described in several ways. Initially, an automated perimeter produces a map of the differential light sensitivities at a range of eccentricities. This is compared with age-adjusted normal sensitivities (total deviation) and described in terms of the likelihood that each point falls within the normal range (total deviation probability plot). The field is then adjusted for overall depression to account for diffuse field loss that may be more likely due to refractive media opacity and thereby to highlight field loss from visual pathway pathology (pattern deviation).⁴ This can also be described as falling within the normal range (pattern deviation probability plot).

Furthermore, the visual field may be described in terms of several numerical values, termed the global indices.^{5–6} These comprise the mean deviation (MD) and the pattern standard deviation (PSD) or the average defect (AD) and the pattern defect (PD) on the HFA and MAP, respectively. The calculation of MD and AD involves averaging the differences between the measured sensitivities and the age-adjusted normal sensitivities

(total deviations) at each test point, thereby describing the general depression or elevation of the field. The PSD and PD describe the spread of these total deviations and represent the asymmetry of the visual field.

Currently, MAP testing results in different numerical global indices from those obtained by the HFA. These indices are therefore not numerically interchangeable between the two devices and it is not clear what values may be substituted when a conversion is required.^{7–8} This study was designed to directly compare the global indices of HFA and MAP visual fields.

METHODS

A total of 63 patients were recruited selectively from a patient population attending an urban glaucoma clinic. None had diabetes, cataract or corneal or retinal disease, which could affect test results. The inclusion criteria were: a visual acuity of 6/12 or better, five dioptres or less of sphere and three dioptres or less of cylinder in refractive error, no previous intraocular surgery and no other systemic illness or medication which could affect visual fields. Patients had suspected glaucoma, ocular hypertension, open angle glaucoma, or were control subjects. The suspected glaucoma group had a family history of glaucoma or had suspicious discs but no definite structural changes and normal intraocular pressure (IOP <21 mm Hg) and visual fields. Ocular hypertension was diagnosed as IOP >21 mm Hg on at least three occasions with no previous field changes on full threshold HFA testing and no evidence of glaucomatous optic neuropathy. Patients with open angle glaucoma had glaucomatous optic disc changes with or without characteristic visual field abnormality on 24-2 HFA full threshold testing. Among patients with suspected glaucoma, ocular hypertension or glaucoma, only those patients were included who had at least two consecutive visual field tests performed within the past 2 years.

One eye from each patient was considered. When both eyes were eligible a random choice was made. Medmont and HFA tests were performed in random order after informed consent was obtained. The Humphrey Field Analyzer II (Carl Zeiss, Dublin, CA, USA) was used to perform central 24-2 full threshold visual field tests. The Medmont M600 Automated Perimeter (Medmont, Camberwell, Victoria, Australia) was used to perform central 30° threshold tests. The global indices from each test were then compared. These included the MD and PSD from the HFA and the AD and PD from the MAP.

Statistical Analysis System 6.12 (SAS Institute, Cary, NC, USA) was used for statistical analysis including frequency tables, descriptive statistics and linear regression.

Abbreviations: AD, average defect; HFA, Humphrey Field Analyzer II; IOP, intraocular pressure; MAP, Medmont Automated Perimeter; MD, mean deviation; PD, pattern defect; PSD, pattern standard deviation




Figure 1 Relationship between Medmont average defect (AD) and Humphrey mean deviation (MD) in differential light sensitivity decibels (dB).

RESULTS

Our sample included 34 females (54%) and 29 males (46%). Their average age was 60 years (standard deviation 13 years). There were 15 controls (24%), eight patients with suspected glaucoma (13%), eight with ocular hypertension (13%) and 32 with open angle glaucoma (51%).⁹

When MD and AD were compared, there was an association which could be described by the quadratic equation:

$$AD = 0.94 + 1.31(MD) + 0.02(MD)^2$$

This regression function was highly significant ($r^2 = 0.92$; $p < 0.001$) (fig 1, table 1).

In addition, when PSD and PD were compared, again there was a non-linear association, which fit the regression equation:

$$PD = 2.21(PSD) - 0.05(PSD)^2 - 0.006$$

This relationship was also highly significant ($r^2 = 0.75$; $p < 0.001$) (fig 2, table 2).

These associations were independent of age and gender.

DISCUSSION

The pattern and extent of visual field loss may indicate the nature and severity of disease, respectively.¹⁰ Comparisons between Humphrey and Medmont perimeters have indicated that visual field loss becomes apparent at the same rate on both devices and demonstrates a similar pattern.^{9 11} Therefore, it might be possible to use the two units interchangeably for disease detection.

The results of this study demonstrate that both MD and PD, and thus disease severity, are comparable with AD and PD, respectively, with the relationships definable by polynomial equations. This complex association may be related to the different spectra of the test stimuli and the number of stimuli, or the different levels of background illuminance between the two devices and thus the difference in absolute sensitivity across the visual field under these different conditions. Both devices use a hemispheric bowl with a radius of approximately 30 cm; however, whilst the HFA illuminates the background to 31.5 apostilbs (asb), the Medmont uses 10 asb.¹² Both use a testing stimulus 0.43° (Goldmann size III) in size. However, the HFA projects a white light stimulus at specified points in the visual field (54 testing locations in the 24-2 algorithm), while the MAP utilises 164 green light-emitting diodes (wavelength 565 nm) which act as test stimuli by retro-illuminating fixed points within the bowl.¹² The larger number of stimuli allows for a greater number of, and thus a greater weighting for, points in the arcuate zones, these being the areas where glaucomatous field loss is more likely to occur initially. Furthermore, the

Table 1 Comparison of values for Humphrey Field Analyzer II (HFA) mean deviation (MD) and Medmont Automated Perimeter (MAP) average defect (AD)

HFA MD (dB)	MAP AD (dB)
-1.00	-0.34
-2.00	-1.59
-3.00	-2.80
-4.00	-3.96
-5.00	-5.09
-8.00	-8.21
-10.00	-10.09
-15.00	-14.05
-20.00	-16.98

luminance of the testing stimuli are scaled at different levels, with responses to MAP stimuli being recorded 10 dB lower than those of HFA.¹³ These differences in stimulus and background luminance do not affect the Weber fraction and thus the ability of the MAP to detect field loss seen on HFA.^{9 11} However, despite this, after adjustment for the difference in background luminance, absolute sensitivities have been measured at approximately 5 dB lower on the MAP when compared with the HFA.¹³

The relationship between the global indices of the devices may also relate to differences in the manner in which they were calculated.¹⁴ For instance, AD is the trimmed mean of the total deviations at each point on the visual field, with the percentage of trimming being determined by the degree of abnormality exhibited in the field (severity of disease, presence of diffuse loss, presence of high false positives).¹⁵ Therefore, fields with large numbers of significantly abnormal points would be more likely to have these points excluded and therefore the AD would be less influenced by them. However, MD is an index which is weighted for eccentricity, with each total deviation amount being divided by the variance of normal values at each point.¹⁴ Since variance tends to increase with eccentricity, peripheral points will contribute less to the calculation of MD than central points.

It is not unexpected that these relationships should be independent of age, as testing point sensitivities are adjusted




Figure 2 Relationship between the Medmont Automated Perimeter pattern defect (PD) and the Humphrey Field Analyzer II pattern standard deviation (PSD) in differential light sensitivity decibels (dB).

Table 2 Comparison of values for Humphrey Field Analyzer II (HFA) pattern standard deviation (PSD) and Medmont Automated Perimeter (MAP) pattern defect (PD)

HFA PSD (dB)	MAP PD (dB)
1.00	2.16
1.20	2.57
1.40	2.99
1.60	3.40
1.80	3.81
2.00	4.21
2.20	4.61
2.40	5.00
2.60	5.39
3.00	6.19
4.00	8.00
5.00	9.74
10.00	16.83

for age prior to the calculation of global indices. However, these relationships are only specific for comparisons between global indices and do not relate to sensitivities at individual testing points.

The sample size was based on the numbers which could be collected by one researcher (AS) during the period of the study and was not pre-determined prior to commencement. However, being clinic-based, our sample comprised a group of patients which should reflect those seen in clinical practice.

The AD and the PD results obtained from the Medmont perimeter may be substituted for the MD and PSD of the Humphrey perimeter after appropriate conversion has been performed.

ACKNOWLEDGEMENTS

We would like to thank Associate Professor Algis Vingrys and the staff at Eye Associates for their support and assistance.

Authors' affiliations

John Landers, Alok Sharma, Ivan Goldberg, Stuart Graham, Eye Associates, Park House, Macquarie Street, Sydney, Australia

John Landers, Royal Adelaide Hospital, Adelaide, Australia

Alok Sharma, Riverina Eye Care Centre, Wagga Wagga, Australia

Ivan Goldberg, Department of Ophthalmology, University of Sydney, Sydney, Australia

Ivan Goldberg, Glaucoma Services, Sydney Eye Hospital, Sydney, Australia

Stuart Graham, Save Sight Institute, University of Sydney, Sydney, Australia

Financial support: This study was not subject to funding from any external source. The authors have no proprietary or financial interest in any of the equipment used in this study.

Performed at Eye Associates, Park House, Macquarie Street, Sydney, Australia

Ethics approval: Ethics approval for this study was obtained from the South Eastern Sydney Area Health Services Clinical Research Ethics committee.

Competing interests: None.

Correspondence to: Dr John Landers, Park House, Floor 4, Suite 2, 187 Macquarie Street, Sydney, NSW 2000, Australia; john.landiers@bigpond.com

Accepted 4 March 2007

Published Online First 27 March 2007

REFERENCES

- 1 Heijl A. Automatic perimetry in glaucoma visual screening; a clinical study. *Graefes Arch Clin Exp Ophthalmol* 1976;**200**:21–37.
- 2 Johnson CA, Keltner JL. Automated suprathreshold static perimetry. *Am J Ophthalmol* 1980;**89**:731–41.
- 3 Li SG, Spaeth GL, Scimeca HA, *et al*. Clinical experiences with the use of an automated perimeter (Octopus) in the diagnosis and management of patients with glaucoma and neurologic diseases. *Ophthalmology* 1979;**86**:1302–12.
- 4 Koucheiki B, Nouri-Mahdavi K, Patel G, *et al*. Visual field changes after cataract extraction: the AGIS experience. *Am J Ophthalmol* 2004;**138**:1022–8.
- 5 Flammer J. The concept of visual field indices. *Graefes Arch Clin Exp Ophthalmol* 1986;**224**:389–92.
- 6 Flammer J, Drance SM, Augustiny L, *et al*. Quantification of glaucomatous visual field defects with automated perimetry. *Invest Ophthalmol Vis Sci* 1985;**26**:176–81.
- 7 Gordon MO, Beiser JA, Brandt JD, *et al*. The Ocular Hypertensive Treatment Study: baseline factors that predict the onset of primary open-angle glaucoma. *Arch Ophthalmol* 2002;**120**:714–20.
- 8 Medeiros FA, Weinreb RN, Sample PA, *et al*. Validation of a predictive model to estimate the risk of conversion from ocular hypertension to glaucoma. *Arch Ophthalmol* 2005;**123**:1351–60.
- 9 Landers J, Sharma A, Goldberg I, *et al*. A comparison of perimetric results with Medmont and Humphrey perimeters. *Br J Ophthalmol* 2003;**87**:690–4.
- 10 Miller NR. Topical diagnosis of lesions in the visual sensory pathway. In: Miller NR, eds. *Walsh and Hoyt's clinical neuro-ophthalmology*, Vol 1, 4th edn. Baltimore: Williams & Wilkins, 1982:108–52.
- 11 Zhang L, Drance SM, Douglas GR. The ability of Medmont M600 automated perimetry to detect threats to fixation. *J Glaucoma* 1997;**6**:259–62.
- 12 Vingrys AJ, Helfrich KA. The Opticom M-600: a new LED automated perimeter. *Clin Exp Optom* 1990;**73**:3–17.
- 13 Pye D, Herse P, Nguyen H, *et al*. Conversion factor for comparison of data from Humphrey and Medmont automated perimeters. *Clin Exp Optom* 1999;**82**:11–13.
- 14 Funkhouser A, Fankhauser F. The effects of weighting the "mean defect" visual field index according to threshold variability in the central and midperipheral visual field. *Graefes Arch Clin Exp Ophthalmol* 1991;**229**:228–31.
- 15 Vingrys AJ, Zele AJ. Robust indices of clinical data: meaningless means. *Invest Ophthalmol Vis Sci* 2005;**46**:4353–7.