EXTENDED REPORT

A comparison of four methods of tonometry: method agreement and interobserver variability

P-A Tonnu, T Ho, K Sharma, E White, C Bunce, D Garway-Heath

Br J Ophthalmol 2005;89:847–850. doi: 10.1136/bjo.2004.056614

See end of article for authors' affiliations

Correspondence to:
D F Garway-Heath, MD,
FRCOphth, Glaucoma
Research Unit, Moorfields
Eye Hospital, London
EC1V 2PD, UK;
david.garway-heath@
moorfields.nhs.uk

Accepted for publication
12 November 2004

Aim: To compare the inter-method agreement in intraocular pressure (IOP) measurements made with four different tonometric methods.

Methods: IOP was measured with the Goldmann applanation tonometer (GAT), Tono-Pen XL, ocular blood flow tonograph (OBF), and Canon TX-10 non-contact tonometer (NCT) in a randomised order in one eye of each of 105 patients with ocular hypertension or glaucoma. Three measurements were made with each method, and by each of two independent GAT observers. GAT interobserver and tonometer inter-method agreement was assessed by the Bland-Altman method. The outcome measures were 95% limits of agreement for IOP measurements between GAT observers and between tonometric methods, and 95% confidence intervals for intra-session repeated measurements.

Results: The mean differences (bias) in IOP measurements were 0.4 mm Hg between GAT observers, and 0.6 mm Hg, 0.1 mm Hg, and 0.7 mm Hg between GAT and Tono-Pen, OBF, and NCT, respectively. The 95% limits of agreement were smallest (bias ± 2.6 mm Hg) between GAT observers, and larger for agreement between the GAT and the Tono-Pen, OBF, and NCT (bias ± 6.7 , ± 5.5 , and ± 4.8 mm Hg, respectively). The OBF and NCT significantly underestimated GAT measurements at lower IOP and overestimated these at higher IOP. The repeatability coefficients for intra-session repeated measurement for each method were ± 2.2 mm Hg and ± 2.5 mm Hg for the GAT, ± 4.3 mm Hg for the Tono-Pen, ± 3.7 mm Hg for the OBF, and ± 3.2 mm Hg for the NCT.

Conclusions: There was good interobserver agreement with the GAT and moderate agreement between the NCT and GAT. The differences between the GAT and OBF and between the GAT and Tono-Pen probably preclude the OBF and Tono-Pen from routine clinical use as objective methods to measure IOP in normal adult eyes.

Intraocular pressure (IOP) measurement has an important role in case detection and management of primary open angle glaucoma. Ocular hypertension (OHT) is associated with an increased risk of developing glaucoma,¹ and reducing IOP has been shown to lessen progressive loss of the visual field.² Accurate and precise measurement of IOP is, therefore, fundamental to management of glaucoma.

Applanation tonometry is the method of measuring IOP with instruments that indent or flatten the corneal apex. The Goldmann applanation tonometer (GAT) is regarded as the "gold standard." However, there are other objective instruments such as the Tono-Pen, ocular blood flow tonograph (OBF), and the non-contact tonometers (NCTs). The NCT has the potential advantage that it uses an air puff to indent the cornea, reducing the possible risk of cross infection with agents such as adenovirus and variant Creutzfeld-Jakob disease.^{3,4} In the United Kingdom, most referrals for suspect glaucoma from the primary care setting (optometry practices) are on the basis of NCT measurements. It is therefore important to determine whether the NCT is sufficiently accurate and precise.

Various studies have compared one or two of these instruments with the GAT,^{5–12} and most studies have not compared inter-tonometer agreement with GAT interobserver agreement (variation in measured IOP arise from both inter-tonometer and interobserver differences). To our knowledge, a broad comparison of all instruments in the same group of patients has not been published.

This study was designed to assess the relative agreement of four methods of tonometry (GAT, Tono-Pen, OBF, and NCT), and to compare inter-instrument agreement with interobserver agreement of GAT IOP measurements.

MATERIALS AND METHODS

The study was conducted in the glaucoma research unit at Moorfields Eye Hospital (London, UK). One hundred and five untreated patients attending the ocular hypertension, normal tension glaucoma, or glaucoma primary care clinics participated in the study. The study was approved by the Moorfields Eye Hospital institutional review board.

Examination was conducted on either eye (chosen randomly) of each patient. Eyes were anaesthetised with Benoxinate and fluorescein drops (Moorfields Eye Hospital Pharmacy, London, UK). Measurements (GAT, Tono-Pen, OBF, and NCT) were performed in a randomised order, with a recovery of about 2 minutes between methods.¹³ Three readings were taken with each instrument, and with the GAT three readings were recorded by each of two observers. The mean of the three readings was used for comparison between tonometers. All tonometers were calibrated at the start of the study and all IOP measurements (except for the GAT) were made and recorded by a single observer (P-AT).

CCT was measured with the Altair ultrasonic pachymeter (Optikron 2000, Rome, Italy) after tonometric measurements had been performed.

Goldmann applanation tonometry

IOP was measured with the Goldmann Applanation Tonometer (Haag-Streit, Bern, Switzerland) by two observers. Observer 1 was a medical student (P-AT) and observer 2 was any one of seven medical staff (four ophthalmologists

Abbreviations: CCT, central corneal thickness; GAT, Goldmann applanation tonometer; IOP, intraocular pressure; NCT, non-contact tonometer; OBF, ocular blood flow; OHT, ocular hypertension

Table 1 Patient data

	Mean	SD	Range
Age (years)	64.6	13.2	23–87
Sex (% male)	50		
Eye side (% right)	52		
CCT (μ m)	546.5	39.8	436–664
GAT observer 1 IOP (mm Hg)	17.2	4.3	09–32
GAT observer 2 IOP (mm Hg)	17.6	4.9	09–35
Tono-Pen IOP (mm Hg)	16.6	4.4	07–29
OBf IOP (mm Hg)	17.1	5.9	08–34
NCT IOP (mm Hg)	16.4	5.0	07–29

CCT, central corneal thickness; GAT, Goldmann applanation tonometer; OBf, ocular blood flow; NCT, non-contact tonometer.

and three ophthalmic technicians). For the majority of measurements, observer 2 was an ophthalmologist (KS). All observers were trained and validated in tonometry. Measurements were made in a masked fashion: one observer set the dial to a "random 0" (between 5 mm Hg and 10 mm Hg); the other observer applanated, turned the dial to obtain the end point without looking at the dial, and the first observer recorded the pressure. The procedure was repeated with the two observers changing roles.

Tono-Pen tonometry

The Tono-Pen XL (Mentor, Santa Barbara, CA, USA) was calibrated daily. The operator touched the cornea with the pen tip several times until a reading was displayed. Only measurements with a standard error smaller than 5% were accepted. If successive measurements differed by more than 5 mm Hg, the procedure was repeated.

Ocular blood flow tonometry

Measurements with the ocular blood flow tonograph (OBf Labs Ltd, Malmesbury, Wiltshire, UK) were made with the slit lamp mounted probe. A new disposable and calibrated OBf tip was used for each subject. The subject's cornea was applanated for 5–10 seconds while approximately 200 IOP measurements were taken and averaged to give the final digital readout.

Non-contact tonometry

The Canon TX-10 non-contact tonometer (Canon USA Inc, One Canon Plaza, Lake Success, NY, USA) automatically recorded three IOP readings. Anaesthetic drops were administered to subjects randomised to have NCT first, so that examination conditions were equivalent to those who had other tonometric measurements beforehand.

Statistical analyses

Analyses were performed in Microsoft Excel 97 SR-2 (Microsoft Corp, Seattle, WA, USA), MedCalc version 7.2.10

Figure 1 Box plot of inter-method differences. The box represents the interquartile range which contains the 50% of values. The line across the box indicates the median. The whiskers are lines that extend from the box to the highest and lowest values, excluding outliers. Circles indicate outliers.

(Mariakerke, Belgium), and SPSS for Windows version 10.0.0 (SPSS Inc, Chicago, IL, USA).

For randomisation, each method was assigned a number and a table of random permutations indicated the order of the instruments to be used for subject.

The effect of repeated testing on IOP was assessed for each technique as the difference between the first and third measurements.

Bland-Altman plots were constructed for comparisons between methods and between GAT observers. The systematic difference between methods was termed the "bias" and random differences were quantified by the "limits of agreement." Where there was no relation between inter-method or interobserver differences and IOP magnitude, bias was calculated as the mean difference, and 95% limits of agreement computed (provided that the differences followed a normal distribution). Where there was a trend of increasing (or decreasing) inter-method/interobserver difference across the range of IOP, regression was conducted and regression based limits of agreement calculated.

Repeatability coefficients were computed as 2.77 times the within subject standard deviation (wsSD) for repeated measurements by the same tonometric method:

$$wsSD = SD(\text{observation 1} - \text{observation 2})/\sqrt{2}$$

RESULTS

Table 1 lists summary data.

Figure 1 is a summary plot of the median and range of IOP measurement differences between GAT observers and between GAT observer 1 and other tonometers. The median

Table 2 Agreement between tonometry methods: bias and 95% limits of agreement (or regression based equivalents)* either side of bias

Tonometry method	Estimated bias	95% limits of agreement
Tono-Pen (mm Hg)	0.6	Bias $\pm$ 6.5
Goldmann observer 2 (mm Hg)†	1.7–(0.12 $\times$ mean)	Bias $\pm$ 2.6
OBf (mm Hg)†	6.0–(0.35 $\times$ mean)	Bias $\pm$ 5.4
NCT (mm Hg)†	3.6–(0.17 $\times$ mean)	Bias $\pm$ 4.8

OBf, ocular blood flow; NCT, non-contact tonometer.

*GAT observer 1 was the reference for comparison.

†Difference between methods related to mean of methods. Bias and 95% limits of agreement calculated from linear regression of method differences against method mean.

Table 3 Estimates (95% limits of agreement) for differences between GAT observer 1 and GAT observer 2, and between GAT observer 1 and other tonometric methods, at various IOP levels

	IOP (mm Hg)			
	10	15	20	25
Goldmann observer 2	0.5 (−2.1, 3.1)	−0.1 (−2.7, 2.5)	−0.7 (−3.3, 1.9)	−1.3 (−3.9, 1.3)
OFB	2.6 (−2.9, 8.0)	0.8 (−4.6, 6.3)	−0.9 (−6.3, 4.6)	−2.6 (−8.1, 2.8)
NCT	1.9 (−2.9, 6.7)	1 (−3.7, 5.8)	0.2 (−4.6, 5.0)	−0.6 (−5.4, 4.2)
Tono-Pen	0.6 (−5.9, 7.1)	0.6 (−5.9, 7.1)	0.6 (−5.9, 7.1)	0.6 (−5.9, 7.1)

difference between GAT observers, and between GAT observer 1 and other tonometers was small.

Table 2 lists the bias and 95% limits of agreement for comparisons between GAT observer 1 and GAT observer 2 and other tonometers. The interobserver differences for the GAT were small in comparison with inter-instrument differences (fig 1 and table 2). The 95% limits of agreement between other pairs of instrument were all wider than the comparisons with GAT (data not shown).

Observer 2, OFB, and NCT had a slight tendency to overestimate IOP measurements made by observer 1 at high pressures. Table 3 sets out the mean difference and 95% limits of agreement between instruments at four IOP levels.

The repeatability coefficient of the GAT observers (2.2 mm Hg and 2.5 mm Hg) was lower than that of the other tonometers (3.2 mm Hg, 3.7 mm Hg, and 4.3 mm Hg for the NCT, OFB, and Tono-Pen, respectively). Two readings by the same observer will be within the repeatability coefficient for 95% of the subjects.

There was no significance difference between the first and third IOP measurements for GAT observer 1, Tono-Pen, or NCT. For GAT observer 2 and the OFB, the first measurement was larger than the third (difference 0.2 mm Hg, $p = 0.06$ and 0.6 mm Hg, $p = 0.02$, respectively).

DISCUSSION

In this study, one GAT observer was constant and the other was any one of a number of trained staff (ophthalmologist or ophthalmic technician). This methodology reflects the clinical setting where patients have IOP measured by different personnel at each visit. There was good agreement between the observers (tables 1, 2), with 95% limits of agreement consistent with previous studies (± 2.2 to ± 3.1 mm Hg, table 4).^{11–14} Agreement was better between the two GAT observers than between the GAT and other tonometers (table 2).

GAT/Tono-Pen comparisons

The 95% limits of agreement between the GAT observer 1 and Tono-Pen XL (table 2) were consistent with previous reports (table 4). The average measurement difference (0.6–1.0 mm Hg, table 1) was small, and there was no tendency for the difference to vary with the level of the IOP (table 2). This is consistent with data reported by Bafa *et al* and Bandyopadhyay *et al* (table 4).^{9–15} In contrast, a tendency for the Tono-Pen to underestimate IOP at high IOP levels has been reported in studies comparing manometric and Tono-Pen IOP measurements,¹⁶ and in those comparing the GAT with the Tono-Pen^{17–18} or Tono-Pen XL.¹⁰

Table 4 Mean difference and 95% limits of agreement between GAT and other tonometric instruments—summary of findings from this and previous studies. Values (mm Hg) are given as means (95% CI); a positive mean difference indicates that Goldmann values are higher

Study	GAT interobserver differences	Tono-Pen	OFB	NCT	NCT type
This study	−0.4 (2.6)‡	+0.6 (6.5)**	+0.1 (5.4)‡	+0.7 (4.8)‡	Canon TX-10
Thorburn, 1978 ¹⁴	−0.7 (3.1)				
Bandyopadhyay <i>et al</i> , 2002 ¹⁵		−1.0 (4.5)**			
Iester <i>et al</i> , 2001 ¹⁰		+0.2 (6.3)**†			
Bafa <i>et al</i> , 2001 ⁹		−0.6 (8.3)**	−0.7 (7.7)		
Midelfart and Wigors, 1994 ²⁶		+2.4 (4.6)***			
Geyer <i>et al</i> , 1992 ⁵		+3.9 (6.5)*			
Frenkel <i>et al</i> , 1988 ¹⁸		+0.8 (6.1)*†			
Minckler <i>et al</i> , 1987 ²⁷		+1.7 (6.1)*			
Kao <i>et al</i> , 1987 ¹⁷		−0.5 (7.4)*†			
Gunvant <i>et al</i> , 2004 ²⁰			−0.5 (6.3)‡		
Yang <i>et al</i> , 2000 ⁸			+0.26 (4.6)†		
Bhan <i>et al</i> , 2003 ¹⁹			−2.7 (6.6)‡		
Jorge <i>et al</i> , 2003 ²¹				−0.1 (3.3)	Reichert AT550
Jorge <i>et al</i> , 2002 ¹²				−0.1 (2.2)	Reichert AT550
				+0.9 (2.5)	Reichert XpertP
Popovich and Shields, 1997 ⁷				+0.6 (4.5)	Reichert XPert
Hansen, 1995 ⁶				+0.9 (4.9)	Reichert XPert
Kretz and Demailly, 1992 ²³				+0.1 (2.9)	Reichert XPert
Hollo <i>et al</i> , 1992 ²⁸				−3.9 (5.7)	Reichert XPert
Myers <i>et al</i> , 1990 ²⁹				−0.7 (5.3)	Reichert XPert
Parker <i>et al</i> , 2001 ¹¹				−0.5 (2.2)	Keeler Pulsair 3000
Mackie <i>et al</i> , 1996 ²²				−1.0 (7.1)	Keeler Pulsair 2000
	−0.1 (2.2)				

OFB, ocular blood flow; NCT, non-contact tonometer.

*Oculab Tono-Pen; **Mentor Tono-Pen XL; ***Bio-Rad Tono-Pen XL; OFB tonograph (OFB Labs).

†Overall value: tonometer overestimates GAT at low IOPs and underestimates at high IOPs.

‡Overall value: tonometer underestimates GAT at low IOPs and overestimates at high IOPs.

GAT/OBF comparisons

On average, the OBF slightly underestimated IOP measurements by the GAT (table 1), in agreement with Yang *et al.*⁸ In contrast, an overestimation by the OBF was reported by Bafa *et al.*⁹ and Bhan *et al.*¹⁹ The 95% limits of agreement were similar to previous findings (table 4).

The OBF underestimated GAT IOP by 2.6 mmHg at 15 mmHg and overestimated GAT IOP by 2.6 mmHg at 25 mmHg (table 3). This finding is in agreement with those of Bhan *et al.*¹⁹ and Gunvant *et al.*²⁰ but contrary to those of Yang *et al.*⁸

GAT/NCT comparisons

There are no reports on the relative performance of Canon TX-10. However, other NCT instruments have been evaluated (table 4).

The 95% limits of agreement between the GAT and Canon TX-10 correspond well with previous reports for the XPERT NCT, but are not as good as those reported for the Keeler Pulsair 3000,¹¹ and the Reichert AT550 in normal¹² and glaucomatous²¹ eyes (table 4).

The NCT had a tendency to overestimate the GAT at high IOP, and underestimate the GAT at low IOP (table 2). Although this effect is not seen in most reports for GAT/NCT comparisons,^{6 11 12 21 22} Kretz and Demailly²³ reported a marginal effect in the same direction. Factors such as CCT (evaluated in the companion paper) may contribute to relative IOP overestimation at higher measured IOP levels.

The repeatability coefficient, for two readings by the same observer, was better for the GAT than the other tonometers, with values comparable to those reported by Pandav *et al.* (2.6 mmHg).²⁴ The value for the NCT is comparable to Vernon's figure for the Keeler Pulsair 2000 (4.2 mmHg and 3.6 mmHg for right and left eyes, respectively).²⁵

The results indicate that repeated readings with the GAT, both within and between observers, are much more reproducible than any of the more automated forms of tonometry. There is moderate inter-instrument agreement between the NCT and GAT and poor agreement between the Tono-Pen and OBF with GAT.

ACKNOWLEDGEMENTS

The authors thank Haag Streit UK for providing the Canon TX-10 non-contact tonometer.

Authors' affiliations

P-A Tonnu, T Ho, K Sharma, E White, D Garway-Heath, Glaucoma Research Unit, Moorfields Eye Hospital, London EC1V 2PD, UK
C Bunce, Department of Research and Development, Moorfields Eye Hospital, London EC1V 2PD, UK

REFERENCES

- Gordon MO, Beiser JA, Brandt JD, *et al.* The ocular hypertension treatment study: baseline factors that predict the onset of primary open-angle glaucoma. *Arch Ophthalmol* 2002;**120**:714–20.
- Heijl A, Leske MC, Bengtsson B, *et al.* Reduction of intraocular pressure and glaucoma progression: results from the Early Manifest Glaucoma Trial. *Arch Ophthalmol* 2002;**120**:1268–79.
- Walia JS, Chronister CL. Possible iatrogenic transmission of Creutzfeldt-Jakob disease via tonometer tips: a review of the literature. *Optometry* 2001;**72**:649–52.
- Desai SP, Sivakumar S, Fryers PT. Evaluation of a disposable prism for applanation tonometry. *Eye* 2001;**15**:279–82.
- Geyer O, Mayron Y, Loewenstein A, *et al.* Tono-Pen tonometry in normal and in post-keratoplasty eyes. *Br J Ophthalmol* 1992;**76**:538–40.
- Hansen MK. Clinical comparison of the XPERT non-contact tonometer and the conventional Goldmann applanation tonometer. *Acta Ophthalmol Scand* 1995;**73**:176–80.
- Popovich KS, Shields MB. A comparison of intraocular pressure measurements with the XPERT noncontact tonometer and Goldmann applanation tonometry. *J Glaucoma* 1997;**6**:44–6.
- Yang YC, Illango B, Cook A, *et al.* Intraocular pressure and pulse rate measurements by the OBF tonograph—comparison to reference instruments. *Ophthalmic Physiol Opt* 2000;**20**:401–407.
- Bafa M, Lambrinakis I, Dayan M, *et al.* Clinical comparison of the measurement of the IOP with the ocular blood flow tonometer, the Tonopen XL and the Goldmann applanation tonometer. *Acta Ophthalmol Scand* 2001;**79**:15–18.
- Iester M, Mermoud A, Achache F, *et al.* New Tonopen XL: comparison with the Goldmann tonometer. *Eye* 2001;**15**:52–8.
- Parker VA, Herrtage J, Sarkies NJ. Clinical comparison of the Keeler Pulsair 3000 with Goldmann applanation tonometry. *Br J Ophthalmol* 2001;**85**:1303–4.
- Jorge J, Diaz-Rey JA, Gonzalez-Mejome JM, *et al.* Clinical performance of the Reichert AT550: a new non-contact tonometer. *Ophthalmic Physiol Opt* 2002;**22**:560–4.
- Recep OF, Hasiripi H, Vayisoglu E, *et al.* Accurate time interval in repeated tonometry. *Acta Ophthalmol Scand* 1998;**76**:603–5.
- Thorburn W. The accuracy of clinical applanation tonometry. *Acta Ophthalmol (Copenh)* 1978;**56**:1–5.
- Bandyopadhyay M, Raychaudhuri A, Lahiri SK, *et al.* Comparison of Goldmann applanation tonometry with the Tonopen for measuring intraocular pressure in a population-based glaucoma survey in rural West Bengal. *Ophthalmic Epidemiol* 2002;**9**:215–24.
- Foster PJ, Wong JS, Wong E, *et al.* Accuracy of clinical estimates of intraocular pressure in Chinese eyes. *Ophthalmology* 2000;**107**:1816–21.
- Kao SF, Lichter PR, Bergstrom TJ, *et al.* Clinical comparison of the Oculab Tono-Pen to the Goldmann applanation tonometer. *Ophthalmology* 1987;**94**:1541–4.
- Frenkel RE, Hong YJ, Shin DH. Comparison of the Tono-Pen to the Goldmann applanation tonometer. *Arch Ophthalmol* 1988;**106**:750–3.
- Bhan A, Bhargava J, Vernon SA, *et al.* Repeatability of ocular blood flow pneumotonometry. *Ophthalmology* 2003;**110**:1551–4.
- Gunvant P, Baskaran M, Vijaya L, *et al.* Effect of corneal parameters on measurements using the pulsatile ocular blood flow tonograph and Goldmann applanation tonometer. *Br J Ophthalmol* 2004;**88**:518–22.
- Jorge J, Gonzalez-Mejome JM, Diaz-Rey JA, *et al.* Clinical performance of non-contact tonometry by Reichert AT550 in glaucomatous patients. *Ophthalmic Physiol Opt* 2003;**23**:503–6.
- Mackie SW, Jay JL, Ackerley R, *et al.* Clinical comparison of the Keeler Pulsair 2000, American Optical MkII and Goldmann applanation tonometers. *Ophthalmic Physiol Opt* 1996;**16**:171–7.
- Kretz G, Demailly P. X-PERT NCT advanced logic tonometer valuation. *Int Ophthalmol* 1992;**16**:287–90.
- Pandav SS, Sharma A, Gupta A, *et al.* Reliability of proton and goldmann applanation tonometers in normal and postkeratoplasty eyes. *Ophthalmology* 2002;**109**:979–84.
- Vernon SA. Reproducibility with the Keeler Pulsair 2000 non-contact tonometer. *Br J Ophthalmol* 1995;**79**:554–7.
- Midelfart A, Wigers A. Clinical comparison of the ProTon and Tono-Pen tonometers with the Goldmann applanation tonometer. *Br J Ophthalmol* 1994;**78**:895–8.
- Minckler DS, Baerveldt G, Heuer DK, *et al.* Clinical evaluation of the Oculab Tono-Pen. *Am J Ophthalmol* 1987;**104**:168–73.
- Hollo G, Follmann P, Pap G. A clinical evaluation of XPERT NCT (Reichert) for glaucoma screening by optometrists. *Int Ophthalmol* 1992;**16**:291–3.
- Myers KJ, Lalle P, Liwak A, *et al.* XPERT NCT—a clinical evaluation. *J Am Optom Assoc* 1990;**61**:863–9.