

	3 × 3-mm	6 × 6-mm	12 × 12-mm
Perfusion density	AUC: 0.945 (95% CI: 0.886-1.003)	AUC: 0.852 (95% CI: 0.756-0.948)	AUC: 0.817 (95% CI: 0.706-0.927)
Vessel length density	AUC: 0.963 (95% CI: 0.920-1.007)	AUC: 0.918 (95% CI: 0.848-0.988)	AUC: 0.897 (95% CI: 0.815-0.978)
Fractal dimension	AUC: 0.964 (95% CI: 0.921-1.007)	AUC: 0.942 (95% CI: 0.886-0.999)	AUC: 0.929 (95% CI: 0.861-0.996)

Supplemental Table 4. Receiver operating characteristic analyses comparing the accuracy of each swept-source optical coherence tomography angiography image scan size and parameter in predicting the presence of diabetic retinopathy in a cohort of all subjects without diabetic macula edema.

AUC, area under the curve; CI, confidence interval