Rechargeable battery belt

For use with the indirect ophthalmoscope

E. W. G. DAVIES AND W. M. BUCHANAN

Department of Experimental Ophthalmology, Institute of Ophthalmology, University of London

A rechargeable battery is a convenient form of providing power for an indirect ophthalmoscope when conducting a ward round in a retinal unit, and dispenses of the need for connexion to a mains supply at each bedside.

Nickel-cadmium cells are widely used in industry and are admirably suited to the supply of power to ophthalmoscopes because of their almost constant voltage during discharge. The light output is thus reasonably constant. Batteries for supplying ophthalmoscopes are usually designed to be carried in the hand or in a capacious pocket, but these are heavy and bulky, and do not leave both hands free for examination of the patient.

BELT

A battery has been designed in the form of a belt which can be carried unobtrusively and when in use leaves the hand free. Five nickel-cadmium cells are carried in pockets attached to the belt and are arranged to give an even distribution of weight.

ILLUMINATION

A three-way switch is also mounted on the belt giving an "off" position and two levels of illumination. The lower level of illumination is achieved by connecting a half-ohm resistor in series with the lamp. To switch one cell out of the circuit would cause unequal discharge of the cells. This lower level of illumination is approximately 15.5 watts at full charge. The higher level of illumination supplies the 18 watt ophthalmoscope bulb.

CHARGER

A suitable charging unit, able to charge four belts simultaneously, can be placed in a central position in the retinal unit and belts are connected to this when the ward round has been completed (Figure). The charge is completed in about 12 hours at 350 millampères, and at this rate no damage occurs if the belts are inadvertently left on charge. The cells can be recharged several hundred times and, although their capacity will gradually decrease with time, their life is prolonged if complete discharge is avoided.

The nominal voltage of each nickel cadmium cell is 1.3 volts when fully charged and the five cells in series therefore provide 6.5 volts.

The belt will sustain continuous use of the lamp for up to 2 hours; with the more usual intermittent use during a ward round the duration of supply is considerably longer.

FIGURE *Charger with four battery belts plugged in*